00/5000/6000 Blower Door QuickGuide



Step 1: Prepare the building

Refer to the Gauge QuickGuide and become completely familiar with its operation before performing a test.

- □ Fireplaces and stoves must be cold with doors closed (cover ashes).
- □ Close and lock outside doors and windows so they don't open during the test.
- □ Open all interior doors leading to conditioned spaces.
- □ Shut off HVAC, combustion appliances, exhaust fans, dryers, A/C and furnaces. Ensure they will not come on □ Turn gas hot water to Pilot.



Step 2: Install the system

- □ Set up the Door Panel. Refer to your "Door Panel QuickGuide"
- □ Connect the Ethernet style Speed Control Cable to the fan and gauge.
- □ Connect the yellow tube between yellow ports marked "Ref B" on fan and gauge. If the fan has a green port ("Input B"), connect the green tube.
- □ Connect the short red tube to the red gauge port.

Traditional Cloth

- \square Pass the red tube through the Door Panel and toss the end at least 5 feet away from the fan's airstream.
- □ Water in tube will result in erroneous readings.



 $\mathbf{SmartCloth}^{\mathsf{TM}}$

near the front, mid-right

 \square Connect the red tube to

the tube insert found

□ SmartCloth can still be

used like the traditional method if desired.

of the cloth.

□ Install the fan blowing outdoors. Cover the fan. Refer to "Range Configurations" QuickGuide for assistance on selecting proper fan coverage.

□ Connect power cord between fan & a wall outlet.

Power light must be on.





□ Place gauge near fan, or attach gauge to Door Panel.



www.retrotec.com Support: (888)330-1345 International: +1(604)732-0142 Made in Everson, WA USA L



Different Results

Tap the **[Channel B]** key to view a different Result, or tap **[Result to be displayed]** on the **[Settings]** menu.

Flow: CFM	Required by many states. Also available in metric units.						
Flow/Area: CFM/sq ft Flow normalized by area	Normalized leakage area is used in many standards. All common units are available.						
ACH: /h Air changes per hour	Air Changes per hour can be shown directly on the gauge. Enter volume as shown below.						
EfLA4/area: sq in/sq ft Normalized EfLA at 4 Pa	Specialized units such as Effective Leakage Area are also available.						
	chosen that requires an Area] or [Volume] will Home screen.						

Volume 22,000 cu ft Tap **[Area]** or **[Volume]** to change . The area and volume can also be changed from the **[Settings]** menu.

Show leakage area Result

Equivalent Leakage Area (EqLA) describes the leakage area in terms of one large hole in a flat surface.

Tap the **[Channel B]** key, then "EqLA: sq in"

, and select

[Channel A] shows the building pressure and **[Channel B]** "EqLA" shows the combined size of all holes in the building.



Leakage area is not a required result, but is a nice way to visualize the size of the hole in the ducts.

Pressurization test

Turn the fan around to blow air into the house.



Tube configuration is the same as the depressurization test.

Adjust fan with gauge

Connect Speed Control Cable to the fan. Solid green Status light indicates gauge is connected.

Tap **[Set Pressure] [50] [Set]** to get gauge to control to a pressure of 50 Pa.



CAUTION: Doors opening during the test will cause fan speed to increase which will create a pressure pulse when that door is re-closed, disturbing dust, soot, and particulates, causing damage. **STOP** the Fan if a door opens.

Tap [Set Speed] [50] [Set] to set speed to 50%.



When **[Set Speed]** or **[Set Pressure]** is used, the Jog keys become active on the **Home** screen. Use the **[JogUp] [JogDown]** arrow keys to increase or decrease the target by 5 Pa or 1%.



Tap [Stop] to turn the fan off.

Adjust fan speed remotely

Use optional WiFi to control the fan remotely. Refer to "GaugeRemote QuickGuide".



Control fan speed with software

Speed control is handled automatically with FanTestic software, for complete automation.

Air Leskage Test in compliance with ASTM E.779-10 Faultesis Pro 24 - Leonard to Retrotec Support.									Select Building Image					
Equipment (Show details)	Primary Gauge: DM32 S/N: 40	2058 Primary F	an: Retrote	c 1000 5/% 1	23456									
Building and Customer deta	alls (Show details)													
Set 1: Pressurization set (S							_							
Start date 2015-06	-02 Start ti	me 16:01	🙆 Get	Time	Pressur	ization	set							
	Operator location	Inside •						Tempe	erature, init	(a)	indoors	68 17	outdoors	68 17
Enter data manual oose one:														
Capture data auto	matically													
	Bas	eline, initial (Pa)												
											Tin	ne per Baseline	Pressure	۰,
	Induc	ed pressure (Pa)	60	55	50	45	40	35	30	25	20	15	10	8
	+ X Test Fan 1	• • •	a] 150	120	115	110	113	100	91	81	72	60	50	44
											Time per	Induced Pressu	re	0,
	Ba	seline, final (Pa)												
-						Show (Graphs		Temperat	ture, final	indoors	68 19	outdoors	68 19
ſ	Total Corre	cted Flow (CFN)	2609	2229	2223	2217	2338	2205	2132	2039	1960	1819	1709	1619
ľ		Error (%)	7.9%	-6.2%	-4.7%	-3.0%	4.7%	1.4%	1.0%	0.2%	0.6%	-1.2%	0.5%	-0.5%
Height X ter	riperature d'ifference 0	n *F			Calculate					Units		confidence		
Air le	akage coefficient, C _p 1083.4	CFM/Pan +/- 1	14.4 %	_					Value		_	interval		
	Exponent, ny 0.19639	+/-	80.3 %	<u> </u>	llear data se	et _		r flow at STP		CFM at 50		*/-3.5% %	_	
Correlation coefficient, r ² 0.9608		-	Air change rate			/n at 50 Pa		+/-3.5% %	_					
						-		nit floor area		CFM/sq ft a	_	+/-3.5% %	_	
								nclosure area		CEM/sq ft a		+/-3.5% %		

Field check gauge weekly

Check for gauge accuracy, blocked, leaking, or pinched tubes weekly, and anytime results are in question.



Check the gauge with the tubes in the Umbilical.

- □ Set **[Time Average]** to 5 seconds in **[Settings]**.
- □ Tap **[Channel B]** and select"Pressure: Pa".
- □ Connect the red tube between the red and yellow ports.

If readings on Channel A and Channel B are within 2% and don't drop noticeably in 5 seconds, the gauge is functioning correctly.

□ Remove red tubes and repeat with the yellow tube between the blue and green ports.

Check gauge and tubes regularly to eliminate the most common source of errors.



Field check system monthly

- □ Perform a Blower Door test on a room and record the EqLA at 50 Pa.
- \Box Install cardboard in upper part of doorway with a 20 x 20 inch hole cut in it.
- Perform a second Blower Door test on the building, record the EqLA at 50 Pa.
- \Box Subtract the first result from the second result and the value should be 400 sq. in. (+/-10%).

